

Catalog Number A16500
Product Name Alignflow™ Flow Cytometry Alignment Beads for Blue Lasers *2.5 µm*
Appearance red suspension
Medium deionized water, 0.05% Tween® 20, 2 mM sodium azide
Concentration 9.2×10^7 particles/mL
Lot Number 2304483

SONICATE WELL BEFORE USE. STORE AT 2° - 8° C, DO NOT FREEZE

	LOT DATA	SPECIFICATION
CV OF THE MEAN INTENSITY¹		
at FL1	2.7 %	≤4.0 %
at FL2	2.1 %	≤4.0 %
at FL3	2.4 %	≤4.0 %
FLOW CYTOMETRY¹		
Percent Singlets	98%	≥ 90%
CV of Forward Scattering	4%	≤ 5%
FLUORESCENCE		
Emission Maximum	575 nm	570 ± 25 nm
MEAN FLUORESCENCE INTENSITY¹		
at FL1	792	700 ± 120
at FL2	790	700 ± 120
at FL3	742	700 ± 120
TECHNICAL DATA²		
Actual Particle Size	2.67 ± 0.067 µm	n.a.
Density of Polystyrene	1.055 g/cm ³	n.a.
MISCELLANEOUS INFORMATION		
Material Lot Number	1896869	n.a.

1. Measured with a calibrated flow cytometer, using 488 nm excitation.

2. Technical data for the unstained microspheres.



Rachel Smith, Quality Assurance Manager
18-May-2017

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.

Thermo Fisher Scientific Corporation
29851 Willow Creek Road
Eugene, OR 97402-9132
Phone (541) 465-8300 Fax (541) 335-0504

Printed Jan 06, 2021