

## Life Technologies

## **CERTIFICATE OF ANALYSIS**

Catalog Number C10045

Product Name CellMask™ Orange plasma membrane stain \*5 mg/mL solution in DMSO\*

Appearance red solution
Lot Number 2295054

	LOT DATA	SPECIFICATION
Absorption <sup>1,2,3</sup> Optical Density	0.98 at 557 nm	0.97 ± 0.07 at 555 nm ± 3 nm

1. 501x dilution

2. Method: Optical Density

3. Solvent: Methanol

Ramulmitt

Rachel Smith, Quality Assurance Manager

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.