

Life Technologies

CERTIFICATE OF ANALYSIS

Catalog Number C16508

Product Name Cell Sorting Set-up Beads for Blue Lasers

Appearance white suspension

Medium distilled water, 0.05% Tween® 20, 2 mM sodium azide

Concentration 2.1 x 10⁷ particles/mL

Lot Number 2257758

SONICATE WELL BEFORE USE. STORE AT 2° - 8° C, DO NOT FREEZE

	LOT DATA	SPECIFICATION
FLOW CYTOMETRY ¹ Percent Singlets	99%	≥ 90%
CV of Forward Scattering	2%	≤ 5%
Mean Fluorescence Intensity at FL1	409 channel value	488 ± 100 channel value
CV of the Mean Intensity at FL1	3 %	≤ 5 %
FLUORESCENCE Emission Maximum	513 nm	515 ± 10 nm
TECHNICAL DATA ² Actual Particle Size Density of Polystyrene	5.6 ± 0.056 μm 1.06 g/cm ³	n.a.
MISCELLANEOUS INFORMATION Material Lot Number	2170980	n.a.

- 1. Measured with a calibrated FACSCalibur™ flow cytometer (BD Biosciences), using 488 nm excitation.
- 2. Technical data for the unstained microspheres.

Ramelsmitt

Rachel Smith, Quality Assurance Manager 2-Dec-2019

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.