

## **Life Technologies**

## **CERTIFICATE OF ANALYSIS**

Catalog Number C36950

Product Name CountBright™ absolute counting beads \*for flow cytometry\* \*100 tests\*

**Appearance** white suspension

**Medium** distilled water, 0.05% Tween® 20, 2 mM sodium azide

Concentration  $0.50 \times 10^5 \text{ beads/}50 \text{ µL}$ 

**Lot Number** 2770930

## SONICATE WELL BEFORE USE. STORE AT 2° - 8° C, DO NOT FREEZE

	LOT DATA	SPECIFICATION
FLOW CYTOMETRY Mean Fluorescence Intensity	meets specification	25,000 - 125,000
FLOW CYTOMETRY Percent Singlets CV of Forward Scattering	98% 6.4%	≥ 90% ≤ 7.5%
TECHNICAL DATA <sup>1</sup> Actual Particle Size	7.08 ± 0.478 µm	n.a.
Density of Polystyrene Specific Surface Area	1.06 g/cm <sup>3</sup> 7.995 x 10 <sup>7</sup> cm <sup>2</sup> /g	n.a. n.a.
Z2 Counter <sup>2</sup> Bead Concentration CV of Bead Count	0.50 x 10 <sup>5</sup> beads/50 μL 1.9%	0.45 - 0.55 x 10 <sup>5</sup> beads/50 μL ≤4.5%
MISCELLANEOUS INFORMATION Material Lot Number	2722640A	n.a.

1. Technical data for the unstained microspheres.

2. Measured using Z2 Coulter Particle Counter

Zach Luedtke, Quality Assurance Manager

Zuch Luetth

14-Sep-2023

Life Technologies Corporation certifies on the date above that this is an accurate record of the analysis of the subject lot, and that the data conform to the specifications in effect for this product at the time of analysis.