

Life Technologies

CERTIFICATE OF ANALYSIS

Catalog Number C36950

Product Name CountBright™ absolute counting beads *for flow cytometry* *100 tests*

Appearance colorless suspension

Medium distilled water, 0.05% Tween® 20, 2 mM sodium azide

Concentration 0.50×10^5 beads/50 µL

Lot Number 2237596

SONICATE WELL BEFORE USE. STORE AT 2° - 8° C, DO NOT FREEZE

	LOT DATA	SPECIFICATION
FLOW CYTOMETRY Mean Fluorescence Intensity	meets specification	25,000 - 125,000
FLOW CYTOMETRY Percent Singlets	92%	≥ 90%
CV of Forward Scattering	7.2%	≤ 7.5%
TECHNICAL DATA ¹ Actual Particle Size	7.1 ± 0.478 µm	n.a.
Density of Polystyrene	1.06 g/cm ³	n.a.
Specific Surface Area	7.995 x 10 ⁸ cm ² /g	n.a.
Z2 Counter² Bead Concentration	0.50 x 10 ⁵ beads/50 μL	0.45 - 0.55 x 10 ⁵ beads/50 μL
CV of Bead Count	2.5%	≤4.5%
MISCELLANEOUS INFORMATION Material Lot Number	2168856	n.a.

1. Technical data for the unstained microspheres.

2. Measured using Z2 Coulter Particle Counter

Ramillamitt

Rachel Smith, Quality Assurance Manager 4-Mar-2020

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.