

Life Technologies

CERTIFICATE OF ANALYSIS

Catalog Number F13083

Product Name FluoSpheres® polystyrene, 1.0 μm, red (580/605) *for tracer studies* *1.0x10¹⁰ beads/mL*

Appearance red suspension

Medium distilled water, 0.02% thimerosal

Concentration 1.0 x 10¹⁰ particles/mL

Lot Number 2291417

SONICATE WELL BEFORE USE. STORE AT 2° - 8° C, DO NOT FREEZE

	LOT DATA	SPECIFICATION
ASSAY		
Dye Loading ¹	21 µg/mg	≥ 12 µg/mg
FLOW CYTOMETRY	000/	. 050/
Percent Singlets	99%	≥ 85%
CV of Forward Scattering	7%	≤ 15%
MICROSCOPY		
Inspection	meets specification	few or no aggregates detectable after sonication
TECHNICAL DATA ²		
Actual Particle Size	0.99 ± 0.014 μm	n.a
Density of Polystyrene	1.055 g/cm ³	n.a
MISCELLANEOUS INFORMATION Material Lot Number	2180261	n.a.

^{1.} μg dye/mg dry microspheres

Ramulmitt

Rachel Smith, Quality Assurance Manager 2-Jan-2020

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.

^{2.} Technical data for the unstained microspheres.