

**Catalog Number** J24974  
**Product Name** Attune™ wash solution for use with the Attune™ Family of Flow Cytometers  
**Appearance** colorless solution  
**Lot Number** 2400818

	LOT DATA	SPECIFICATION
<b>Assay Test</b>	11.6	pH between 11.3 and 11.8
<b>Assay Test</b>	Pass	appearance should be clear, colorless
<b>MISCELLANEOUS INFORMATION</b> Date of Manufacture	21-Sep-2021	n.a.
<b>MISCELLANEOUS INFORMATION</b> Expiration Date	21-Sep-2023	n.a.



Rachel Smith, Quality Assurance Manager  
21-Sep-2021

*Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*