

**Catalog Number** J24974  
**Product Name** Attune™ wash solution for use with the Attune™ Family of Flow Cytometers  
**Appearance** colorless solution  
**Lot Number** 2509772

	LOT DATA	SPECIFICATION
<b>Assay Test</b>	11.5	pH between 11.3 and 11.8
<b>Assay Test</b>	Pass	appearance should be clear, colorless
<b>MISCELLANEOUS INFORMATION</b> Date of Manufacture	25-Jun-2022	n.a.
<b>MISCELLANEOUS INFORMATION</b> Expiration Date	25-Jun-2024	n.a.



Aaron Rider, Quality Control Supervisor  
 25-Jun-2022

*Life Technologies Corporation certifies on the date above that this is an accurate record of the analysis of the subject lot, and that the data conform to the specifications in effect for this product at the time of analysis.*