CERTIFICATE OF ANALYSIS

Catalog Number P11495

Product Name Quant-iT™ PicoGreen® dsDNA reagent *2000 assays* *10 x 100 μL*

Appearance orange solution

Lot Number 2139327

	LOT DATA	SPECIFICATION
Absorption ¹		
Optical Density	0.38 at 498 nm	0.36 ± 0.04 at 498 nm ± 3 nm

1. Solvent: Buffer, 50 mM potassium phosphate (pH 7). Method: Optical Density. 51x dilution



Rachel Smith, Quality Assurance Manager 20-Aug-2019

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.