

Life Technologies

CERTIFICATE OF ANALYSIS

Catalog Number Q10211

Product Name Qubit™ RNA BR Assay Kit, 500 assays

Lot Number 2186086

	LOT DATA	SPECIFICATION
ABSORPTION ¹ Optical Density A (lambda max)	meets specification	OD ± 0.07 at lambda max ± 5 nm at lambda max
ABSORPTION Optical Density of Component D ²	0.26 at 260 nm	0.25 ± 0.01 at 260 nm

1. Solvent: DMSO. Method: Optical Density: 10X dilution.

2. Solvent: Buffer, TE Buffer. Method: Optical Density. 10X dilution



Rachel Smith, Quality Assurance Manager

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.