

## **Life Technologies**

## **CERTIFICATE OF ANALYSIS**

Catalog Number Q32850

Product Name Qubit™ dsDNA BR Assay Kit, 100 assays

**Lot Number** 2411538

	LOT DATA	SPECIFICATION
ABSORPTION <sup>1</sup> Optical Density A (lambda max)	meets specification	OD ± 0.07 at lambda max ± 5 nm at lambda max
ABSORPTION Optical Density of Component D <sup>2</sup>	0.98 at 260 nm	1.0 ± 0.04 at 260 nm

1. Solvent: Methanol. Method: Optical Density: 50X dilution.

2. Solvent: Buffer, TE Buffer. Method: Optical Density. 2X dilution

Ramusmitt

Rachel Smith, Quality Assurance Manager

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.