

Catalog Number Q32891
Product Name Qubit™ Endotoxin Detection Assay Kit *for use with the Qubit™ Flex Fluorometer*
Lot Number 2736721

	LOT DATA	SPECIFICATION
Correlation Coefficient of Standard Curve <i>r</i>	0.998	≥ 0.980
CV Average of Sample Replicates CV Average	3 %	$\leq 10\%$
Background Separation Z Score	meets specification	≥ 0.0
Qubit Incubation Time Incubation Time	22 minutes	$10 \leq X \leq 40$ minutes
Quant-iT Incubation Time Incubation Time	24 minutes	$10 \leq X \leq 40$ minutes



Zach Luedtke, Quality Assurance Manager
18-Sep-2023

Life Technologies Corporation certifies on the date above that this is an accurate record of the analysis of the subject lot, and that the data conform to the specifications in effect for this product at the time of analysis.