

Catalog Number Q32892
Product Name Quant-iT™ Endotoxin Detection Assay Kit
Lot Number 2706892

	LOT DATA	SPECIFICATION
Correlation Coefficient of Standard Curve <i>r</i>	0.9983	≥ 0.980
CV Average of Sample Replicates CV Average	4 %	≤ 10%
Background Separation Z Score	meets specification	≥ 0.0
Quant-iT Incubation Time Incubation Time	24 minutes	10 ≤ X ≤ 40 minutes



Zach Luedtke, Quality Assurance Manager
3-May-2023

Life Technologies Corporation certifies on the date above that this is an accurate record of the analysis of the subject lot, and that the data conform to the specifications in effect for this product at the time of analysis.