MOLECULAR PROBES®

CERTIFICATE OF ANALYSIS

Catalog Number Q33223

Product Name Qubit™ RNA XR Assay Kit, 100 Assays

Lot Number 1990303

	LOT DATA	SPECIFICATION
ABSORPTION Optical Density A (lambda max)	meets specification	OD ± 0.09 at lambda max ± 3 nm at lambda max
ABSORPTION		
Optical Density of Component D	0.50 at 260 nm	0.50 ± 0.02 at 260 nm



Rachel Smith, Quality Assurance Manager

15-Jun-2018

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.