CERTIFICATE OF ANALYSIS

Catalog Number Q33235

Product Name Qubit™ RNA IQ Assay - RNA Standards

Lot Number 2105400

	LOT DATA	SPECIFICATION
ABSORPTION		
Optical Density of Component A	0.51 at 260 nm	0.50 ± 0.02 at 260 nm
ABSORPTION		
Optical Density of Component B	0.51 at 260 nm	0.50 ± 0.02 at 260 nm



Rachel Smith, Quality Assurance Manager 19-Apr-2019

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.