

Catalog Number Q33235
Product Name Qubit™ RNA IQ Assay - RNA Standards
Lot Number 2352013

	LOT DATA	SPECIFICATION
ABSORPTION Optical Density of Component A	0.50 at 260 nm	0.50 ± 0.02 at 260 nm
ABSORPTION Optical Density of Component B	0.49 at 260 nm	0.50 ± 0.02 at 260 nm



Rachel Smith, Quality Assurance Manager
25-May-2021

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.