

**Catalog Number** Q33237  
**Product Name** Qubit™ 4 Fluorometer System Verification Assay Kit  
**Lot Number** 2157172

	LOT DATA	SPECIFICATION
<b>ABSORPTION</b> Ratio	105 %	85%-115%
<b>ABSORPTION</b> Ratio	86 %	85%-115%
<b>ABSORPTION</b> Signal to Background	675	≥ 50
<b>ABSORPTION</b> Signal to Background	1227	≥ 10



Rachel Smith, Quality Assurance Manager  
23-Oct-2019

*Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*