

**Catalog Number** Q33237  
**Product Name** Qubit™ 4 Fluorometer System Verification Assay Kit  
**Lot Number** 2208533

	LOT DATA	SPECIFICATION
<b>ABSORPTION</b> Ratio	93 %	85%-115%
<b>ABSORPTION</b> Ratio	90 %	85%-115%
<b>ABSORPTION</b> Signal to Background	730	≥ 50
<b>ABSORPTION</b> Signal to Background	1232	≥ 10



Rachel Smith, Quality Assurance Manager  
 20-Mar-2020

*Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*