

**Catalog Number** Q33237  
**Product Name** Qubit™ 4 Fluorometer System Verification Assay Kit  
**Lot Number** 2234964

	LOT DATA	SPECIFICATION
<b>ABSORPTION</b> Ratio	100 %	85%-115%
<b>ABSORPTION</b> Ratio	94 %	85%-115%
<b>ABSORPTION</b> Signal to Background	808	≥ 50
<b>ABSORPTION</b> Signal to Background	1082	≥ 10



Rachel Smith, Quality Assurance Manager  
13-Aug-2020

*Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*