

**Catalog Number** Q33237  
**Product Name** Qubit™ 4 Fluorometer System Verification Assay Kit  
**Lot Number** 2268317

	LOT DATA	SPECIFICATION
<b>ABSORPTION</b> Ratio	110 %	85%-115%
<b>ABSORPTION</b> Ratio	113 %	85%-115%
<b>ABSORPTION</b> Signal to Background	719	≥ 50
<b>ABSORPTION</b> Signal to Background	1029	≥ 10



Rachel Smith, Quality Assurance Manager  
3-Nov-2020

*Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.*