

Catalog Number Q33237
Product Name Qubit™ 4 Fluorometer System Verification Assay Kit
Lot Number 2403683

	LOT DATA	SPECIFICATION
ABSORPTION Ratio	101 %	85%-115%
ABSORPTION Ratio	87 %	85%-115%
ABSORPTION Signal to Background	776	≥ 50
ABSORPTION Signal to Background	1014	≥ 10



Rachel Smith, Quality Assurance Manager
27-Sep-2021

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.