

Life Technologies CERTIFICATE OF ANALYSIS

Catalog Number	S21384	
Product Name	streptavidin, Alexa Fluor® 750 conjugate	
Appearance	blue solid	
Lot Number	2286334	

Small variation in color is possible, but should not affect product performance.

	LOT DATA	SPECIFICATION
ABSORPTION Maximum	755 nm	755 ± 5 nm
Degree of Labeling ¹	3	2 – 3
FLUORESCENCE Emission Maximum	771 nm	772 ± 5 nm
Relative Quantum Yield ²	5.8	≥ 2.5
MICROSCOPY Immunocytochemistry ³	meets specification	good staining, negligible background
PURITY ⁴ TLC	meets specification	negligible or no free dye detected

1. Moles of dye per mole of protein, determined using an ε of 270,000 cm⁻¹M⁻¹ at the absorption maximum

2. Quantum yield determined relative to IR 125

3. Cell staining performed using a primary antibody specific for cell type and isotype, and this product

4. Solvent: 80% acetonitrile/20% water

2an Monitt

Rachel Smith, Quality Assurance Manager 26-Jan-2021

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.

Thermo Fisher Scientific Corporation 29851 Willow Creek Road Eugene, OR 97402-9132 Phone (541) 465-8300 Fax (541) 335-0504 Printed Jan 27, 2021