

Life Technologies

CERTIFICATE OF ANALYSIS

Catalog Number Z25306

Product Name Zenon® Alexa Fluor® 568 Rabbit IgG Labeling Kit *50 labelings*

Lot Number 2174882

	LOT DATA	SPECIFICATION
ABSORPTION ¹ Maximum	577 nm	579 ± 5 nm
Degree of Labeling ²	2	1 – 3
ACTIVITY TEST ¹ HPLC ³	meets specification	formation of complex with rabbit IgG
FLUORESCENCE ¹ Emission Maximum	599 nm	599 ± 5 nm
Relative Quantum Yield ⁴	0.9	≥ 0.5
MICROSCOPY ¹ Immunocytochemistry ⁵	meets specification	good staining, negligible background
PURITY ¹ HPLC ³	meets specification	≥90%
TLC ⁶	meets specification	negligible or no free dye detected

- 1. For Component A alone
- 2. Moles of dye per mole of protein, determined using an ε of 91,300 cm⁻¹M⁻¹ at the absorption maximum
- 3. HPLC performed using a Biosep S-3000 column
- 4. Quantum yield determined relative to sulforhodamine 101
- 5. Cell staining performed using a primary antibody specific for cell type and isotype, and this product
- 6. Solvent: 80% acetonitrile/20% water

Ramulmitt

Rachel Smith, Quality Assurance Manager 14-Jan-2020

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.