thermo scientific

MabCaptureC Affinity Matrix

Origin: United Kingdom

Certificate of Analysis

Catalogue Code:

196366205L

Lot Number:	3344LL0100	
Retest Date:	Aug 2027	
Manufacturing Date:	Aug 2024	
Storage Temperature:	2 / 8 °C	

TEST	SPECIFICATION	RESULT	UNITS
Dynamic Binding Capacity	≥ 50	62	mg/mL, 10% breakthrough, 4.8 min residence time hlgG
Pressure Flow	> 500	751	cm/hr at 3 bar in a 2.6 x 20 cm column
Microscopy	≥ 95	100	% perfect beads
Microbial Contamination	≤ 20	< 1	CFU/mL suspension
Endotoxin	< 5	< 5	EU/mL
Visual Appearance	White to off white slurry	Conforms	

For Research Use or Further Manufacturing. Not for diagnostic use or direct administration into humans or animals.

M. Schlichertin

Electronically signed by: Martin Schlichting Reason: Author of the GxP document Date: Sep 13, 2024 15:20 GMT+2

Quality Department