

Certificate of Analysis

Product No.: 042695

Product: Potassium chloride, 0.0765M, Conductivity Standard

Lot No.: X01K509

 $\begin{array}{lll} \mbox{Appearance} & \mbox{Clear liquid} \\ \mbox{Actual at 25 °C} & \mbox{10,000 } \mu\mbox{S/cm} \\ \mbox{Uncertainty} & \mbox{\pm 0.31 \%$} \end{array}$

Certification and Traceability: Thermo Fisher Scientific is ISO 9001:2015 certified. This CRM was manufactured and certified by a Thermo Fisher Scientific supplier under ISO 9001, ISO/IEC 17025, and ISO 17034 Quality Management System.

Expiration Date: November 1, 2025

Order our products online thermofisher.com/chemicals

This document has been electronically generated and does not require a signature.