

Certificate of Analysis

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
Standard ISO9001:2015 by LRQA. Certificate Number. 10551915

Product No.:	A17734/L19566
LOT Number:	10251132
CAS Number:	9002-26-0
Description	Amberlite IRA-410(Cl), ion exchange resin
Quality Test/Release Date:	05/06/2024
Retest Date:	Jun/2029
Country of Origin:	France

Result Name	Test Value
Appearance:	Pale yellow beads
Particle size:	0.1% <300um
Particle size:	0.0% >1180um
Water retention capacity:	48.2%
Harmonic mean size:	0.666mm
Total exchange capacity:	1.38eq/L
Uniformity Coefficient:	1.47



Date printed: 5 June, 2024
Jonathan Davis, Quality Control Team Leader