

## Certificate of Analysis

Product No.: J60565

Product: Sodium phosphate, 0.2M buffer soln., pH 9.0

Lot No.: S23I731

Appearance Clear liquid

pH  $9.0 \pm 0.15$ 

NaH<sub>2</sub>PO<sub>4</sub>/Na<sub>2</sub>HPO<sub>4</sub> 200 mM

Method: Ultrapure sodium phosphates in given concentration was dissolved in 18.2 megohm-cm ± 1 water (CDI, 18 megohm, UV treated), adjusted to pH 9.0 and then filtered through 0.22- micron filter.

Expiration Date: June 15, 2025

Order our products online thermofisher.com/chemicals

This document has been electronically generated and does not require a signature.

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.