

**BECKMAN COULTER SYSTEMS
DRUGS OF ABUSE ASSAYS (DAT)
CERTIFICATE OF ANALYSIS**

MICROGENICS PART NUMBER	1604
BECKMAN COULTER PART NUMBER	474985
MICROGENICS LOT NUMBER	74281965

A. CALIBRATOR/MANUFACTURE INFORMATION

PRODUCT DESCRIPTION	OPIATE 300 HIGH CALIBRATOR (1000 NG/ML)
MICROGENICS PART NUMBER	1604
BECKMAN COULTER PART NUMBER	474985
VENDOR (MGC) LOT NUMBER	74281965
PLATFORM USED FOR MANUFACTURE TESTING	DXC
MANUFACTURE DATE	April 2021
EXPIRATION DATE	2022-10-31
QUANTITY	799
STORAGE CONDITION	2° TO 8° C

B. MATERIALS USED FOR TESTING

		BCI P/N	MGC P/N	LOT NUMBER	EXPIRATION
1. CALIBRATORS FROM MOST RECENTLY APPROVED LOT					
NEGATIVE CALIBRATOR		445803	0555	73872060	2021-09-30
OPIATE 300 LOW CALIBRATOR (300 ng/mL)		474982	1602	73852189	2021-07-31
OPIATE 300 HIGH CALIBRATOR (1000 ng/mL)		474985	1604	73871932	2021-08-31
2. GOLD STANDARDS					
OPIATE 1000 ng/mL GOLD STANDARD		NA	2069	72799447	2021-08-28
3. REAGENTS					
OPIATE 300	ANTIBODY	445963	10007370	73777181	2021-07-31
	ENZYME	445962	0550	73777182	2021-07-31

C. TESTING*

	SPECIFICATION	RESULT	PASS/FAIL
1. FILL VOLUME	5.2 ± 0.2 mL	5.2 mL	PASS
2. APPEARANCE	YELLOWISH CLEAR SOLUTION; ESSENTIALLY FREE OF PARTICULATE	Yellowish Clear Solution; Essentially Free Of Particulate	PASS
3. UPN/GS1 BARCODE	Meets SCD specifications (Yes or No)	Yes	PASS
4. ANALYTE CONCENTRATION (BY GC/MS)			
OPIATE 1000 ng/mL	900 ng/mL TO 1100 ng/mL	986 ng/mL	PASS

**BECKMAN COULTER SYSTEMS
DRUGS OF ABUSE ASSAYS (DAT)
CERTIFICATE OF ANALYSIS**

MICROGENICS PART NUMBER	1604
BECKMAN COULTER PART NUMBER	474985
MICROGENICS LOT NUMBER	74281965

C. Testing (Continued)*

5. COMPARISON TO GOLD STANDARD AND MOST RECENTLY APPROVED LOT ON A SYNCHRON SYSTEM AT 37 °C USING BECKMAN DEFINED PARAMETERS:

SPECIFICATION:	THE RATE IN mA/min MUST BE WITHIN ± 5% OF GOLD STANDARD AND MOST RECENTLY APPROVED LOT .					
	SAMPLE RESULT	GOLD STANDARD		RECENT LOT		PASS/FAIL
		RESULT	PERCENT	RESULT	PERCENT	
OPIATE 300	435	433	0%	435	0%	PASS

D. TREND ANALYSIS DATA

DATA ATTACHED	BY	DATE
YES	Magiruz Hilario	06/03/2021

E. COMMENTS:

* Testing performed on a Synchron Automated Chemistry Analyzer at 37 °C.

TESTED BY: Magiruz Hilario

DATE: 06/03/2021

APPROVED BY: [Signature]

DATE: 06/07/2021

QUALITY CONTROL

Opiate 300 High Calibrator (1000 ng/mL)

BCI P/N: 474985

MGC P/N: 1604

Lot Numbers	MGC Gold Standard	Reference	GC/MS Results
56467304	-1	0	1016
56996452	-2	-1	978
57011168	-1	0	978
57342084	-2	0	978
57488801	-1	0	978
57806718	-1	1	980
58031778	-1	0	980
58192830	-1	0	980
58332630	-1	0	980
58499305	-1	0	980
58700526	-1	0	980
58787623	0	0	980
58857840	-1	0	1084
58950406	-1	0	1084
59038811	0	0	1084
59088714	-1	0	1084
59182675	-1	0	1084
59283209	0	0	983
59351314	-1	-1	1084
59521450	0	1	983
59488443	-1	0	983
59579364	2	3	983
59633982	0	-2	1024
59716381	-2	-2	1024
59784395	1	0	1024
59872088	0	0	1024
59977150	0	0	1024
60176978	0	0	1024
60247748	0	1	1024
60504755	0	0	1024
60550499	0	0	1091
60665810	0	0	1091
70782840	0	0	1091
72092041	0	0	1091
72159394	0	0	1091
72197144	0	0	1091
72409579	-1	-1	1006
72629071	0	-1	1006
72635613	0	0	1006
72712056	0	0	1006
72851997	0	0	1006
73082810	1	1	969
73187085	0	0	969
73310421	1	0	1020
73453428	0	-1	1020
73623100	0	0	1020
73871932	0	0	1020
74088216	0	0	986
74281965	0	0	986

