

Certificate of Analysis

Product Name: Cholesterol,Infinity,2x125 mL
Product Numbers: TR13421
Lot Number: 200031

Date of Manufacture: 24-FEB-25
Expiration Date: 2026-08-31
Storage: 2 - 8°C

Specification	Specification Limit	Result
Appearance	Clear, colorless to slightly pink solution, essentially free of particulate matter	Complies
Data-Trol Normal	Lower Limit	3.53 mmol/L
Data-Trol Normal	Upper Limit	4.59 mmol/L
Data-Trol Normal	Result	4.34 mmol/L
Data-Trol Normal	Lot Number	1550UN
Data-Trol Abnormal	Lower Limit	6.31 mmol/L
Data-Trol Abnormal	Upper Limit	8.19 mmol/L
Data-Trol Abnormal	Result	7.72 mmol/L
Data-Trol Abnormal	Lot Number	1212UE
Linearity	Linear to 20 mMol/L	Complies

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications and has been released.

This product contains animal source material. The bovine material used in the manufacture of this product is derived from bovine blood collected from U. S. sourced cattle slaughtered at a USDA licensed establishment.

Tiffany S Echoff

Quality Representative

Date: 2025-04-04