

Certificate of analysis

N,N-Dimethyl-1,4-butanediamine

Stock Number: H50276

Lot Number: K14S015

Analysis

Assay 97+ % by VPC
Column: DB-1 (100% -polydimethylsiloxane),
Length: 30 m, I.D.: 0.25 mm, Film: 1.0 µl
Injection temp: 250°C
Detector (FID) temp: 275°C
Column temp: 100°C initial, programmed at
10°C/min to 250°C, hold for 15 min
Sample: 0.2 µl neat

Appearance Clear, colorless liquid

Certified by:



Quality Control

This document has been electronically generated and does not require a signature.