

# Certificate of analysis

**1-Chloro-2-iodoethane, 96%**

**Stock Number: L10742**


**Lot Number: J27W029**

## Analysis

Assay 98.5 % by GC  
Column: DB-1 (100% -polydimethylsiloxane),  
Length: 30 m, I.D.: 0.25 mm, Film: 1.0 µl  
Injection temp: 250°C  
Detector (FID) temp: 275°C  
Column temp: 40°C initial, programmed at  
10°C/min to 250°C, hold for 15 min  
Sample: 0.2 µl neat

Appearance Clear, colorless liquid

Certified by:



Quality Control

This document has been electronically generated and does not require a signature.