

Certificate of analysis

3-Bromo-4-(trifluoromethoxy)aniline

Stock Number: H26174

Lot Number: D11T004

Analysis

Assay 99.8 % by VPC
Column: DB-1 (100% -polydimethylsiloxane),
Length: 30 m, I.D.: 0.25 mm, Film: 1.0 µl
Injection temp: 250°C
Detector (FID) temp: 275°C
Column temp: 80°C initial, programmed at
10°C/min to 250°C, hold for 15 min
Sample: 0.2 µl neat

Appearance Clear, very pale yellow liquid

Certified by:



Quality Control

This document has been electronically generated and does not require a signature.