

# Certificate of analysis

## 5-Fluoro-2-(trifluoromethyl)benzylamine

**Stock Number: H26751**


**Lot Number: K27S043**

### Analysis

Assay 97% (96.5 %) by VPC  
Column: DB-1 (100% -polydimethylsiloxane),  
Length: 30 m, I.D.: 0.25 mm, Film: 1.0 µl  
Injection temp: 250°C  
Detector (FID) temp: 275°C  
Column temp: 60°C initial, programmed at  
10°C/min to 250°C, hold for 15 min  
Sample: 0.2 µl neat

Appearance Clear, light yellow liquid

Certified by:



Quality Control

This document has been electronically generated and does not require a signature.