Certificate of analysis

Product No.: J63379

Product: TES, 0.2M buffer soln., pH 7.0

Lot No.: X12E572

Appearance Clear liquid

pH 7.0 ± 0.05

TES 0.2 M

Method: Ultrapure TES in given concentration was dissolved in 18.2 megohm-cm ± 1 water (CDI, 18 megohm, UV treated), adjusted to pH, and then filtered through sub-micron filter.

This document has been electronically generated and does not require a signature.

