



422 Aldo Avenue Santa Clara, CA 95054
75 Panorama Creek Drive, Rochester, NY 14625

Product Certificate

Thermo Scientific
Nalgene and Fiberlite Products

This information is the product certification for item # **010-1494**

Description:

Thermo Scientific Fiberlite Bottle 500 mL PC (sets of 6)

Lot#:

—

Manufactured:

8/4/2021

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI
010-1394	Thermo Scientific Fiberlite Bottle 500 mL	Component Part	-	-	-
	Resin used for Bottle	Polycarbonate	-	Passed	Passed
010-1473	Thermo Scientific Fiberlite Bottle 500 mL Closure	Component Part	-	-	-
	Resin used for closure	PP, 20% glass filled	-	Passed	Passed
010-1474	Thermo Scientific Fiberlite Bottle 500 mL Closure liner	Component Part	-	-	-
	Resin used for closure liner	Polypropylene	-	Passed	Passed
001-0299	Thermo Scientific Fiberlite Bottle 500 mL O-ring	Component Part	-	-	-
	Resin used for O-ring	Silicone	-	Passed	Passed

Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity and USP Class VI.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol. If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol. The closure and plug and the silicone o-ring have been tested and shown to comply with USP Class VI requirements and to non-cytotoxic. The polypropylene/polycarbonate resin for the bottle meets the requirements of 21 CFR section 177.1520.