

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 Quality Management System (CERT-0040858 by QMI-SAI Global) in the USA.

 Robert Prescott  
 Mgr. QA/RA

The following information represents Product Certification for: Item#: **312084-9025**

Certificate issued: **07/24/2009**

Description: **BTL PKG TRANS AMBER HDPE;1/4OZ**

Lot#: **1004700**

Manufactured: **07/02/2009**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0435-61P	BTL,8ML,RND,N/M,HDPE,TRAN AMB	COMPONENT PART				
8-0042-34P	RESIN,HDPE,TRSL AMB,IBM,EBM	COLOR MIX (HDPE, TRSL AMB)	N/A	PASSED	N/A	N/A
8-0042-31	RESIN,HDPE,IBM,EBM,EXT	HIGH-DENSITY POLYETHYLENE	3310	PASSED	PASSED	177.1520 (c) 3.2a
8-0099-16	COLOR,TRSL AMB,OLF	COLORANT, TRSL. AMBER	20858	N/A	N/A	177.1520, 178.2010, 3297
1-1811-62	CLOS,20/415,PP,AMB,NALGE	COMPONENT PART				
8-0070-44P	RESIN,PP,AMB,INJ	COMPONENT PART				
8-0070-02	RESIN,PP,INJ	POLYPROPYLENE, INJECTION	12204	PASSED	PASSED	177.1520(a)(1)(i), (c)1.1a(use conditions A-H)
8-0099-51	COLOR,AMBER, MULTI,	COLORANT, OPAQUE, AMBER	20840	PASSED	PASSED	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.