
Certificate of Conformity

Product: Nucleolink Strips 8 wells, Clear**Date:** 2022-02-16**Page:** 1**Item number:** 248259**Lot number:** 173838**Date of manufacture:** January 2022**Date of expiry:** January 2027

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements" and ISO 13485:2016 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site

Thomas Andersen
Operations Quality Manager

Please note that this is an electronically generated signature