

---

## Certificate of Conformity

**Product:** Nunc Snap Caps for Nunc Container, Natural

Date: 2023-10-03

Page: 1

**Item number:** 536617

**Lot number:** 180302

**Date of manufacture:** September 2023

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP <88> "Biological Reactivity test In Vivo, USP Class VI Test"	USP current edition	Complies**

\*\*Only performed once

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements" and ISO 13485:2016 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site



Anita Kvistgaard  
Operations Quality Manager

Please note that this is an electronically generated signature