

## Certificate of Conformity

Date: 23 Nov 2020

**Product:** 350ml Bottle Dosed

**Item Number:** WSC350

**Lot Number:** 350/12

**Date of Expiry:** 05/2022

Thermo Fisher Scientific confirms that the product supplied has been manufactured in a controlled environment, inspected and tested in accordance to relevant specifications.

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10<sup>-6</sup>.

	Min (kGy)	Max (kGy)	Certificate of Irradiation	Date of Irradiation
Calculated Dose	22.7	28.0	23,511	23/24 Jun 20
Specification	15.0	40.0		

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2015 "Quality management system - Requirements".

Signed on behalf of the Quality Department:

*J. Mallett*

Batch Release-Document Controller