

Certificate of Analysis

Date: 05/29/2020

Catalog Number: 411G10R27

Lot Date: 200301

Product Description: 4OZ HP W/PILL/STRING

Component Material: Virgin Polypropylene

Attribute: 120 ml

This product has been sampled in accordance with Standard Operating Procedure and has been release by the Quality Assurance for the following Characteristics.

Test / Procedure Results

Visual Attributes: PASS

Leak Tested: PASS

Sterility: Has passed the required sterility certification that we have implemented as per ANS/AAMI/ISO 11137 Part 1 and 2. This is done by a third party.

Sodium Thiosulfate Tablet LOT #: 110518A

Sodium Thiosulfate (as pentahydrate) (%): 17.8

Sodium Thiosulfate Tablet Expiration: 11/05/2022

Fluorescence Verification (UV Lamp – 365NM (6 watt bulb): PASS

Volume Accuracy: Tolerance of this container to the 100 ml Fill Line is +/-2.5 ml

Expiration Date Sterility: The Expiration Date of Sterility is two years from the LOT Date.



Jerry Acoff
Quality Engineering Manager