

Bottle Type: Glass

QA Level: 1

Lot Number: 00085650

Item Number: 02912313

Item Description: 1000mL WM Clear SS Jar



Groups 1 and 2 are applicable

Fisherbrand Products Level 1 products have been tested and found to comply with or to be lower than the EPA detection limits as stated in OSWER Directive # 9240.005A “Specifications And Guidance For Contaminant-Free Sample Containers 12/92”. Representative containers have been tested at periodic intervals by certified third party laboratories in accordance with our quality procedures.

Group 1. Glass and HDPE Sample containers for use in the analysis of Metals

Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)
Aluminum	<80	Calcium (all HDPE)	<100	Magnesium	<100	Selenium	<2
Antimony	<5	Chromium	<10	Manganese	<10	Silver	<5
Arsenic	<2	Cobalt	<10	Mercury	<0.2	Sodium	<500
Barium	<20	Copper	<5	Nickel	<10	Sodium (all HDPE)	<100
Beryllium	<0.5	Iron	<50	Potassium	<750	Thallium	<5
Cadmium	<1	Lead	<2	Potassium (all HDPE)	<100	Vanadium	<10
Calcium	<500					Zinc	<10

In addition to the above analytes, HDPE containers are certified for these analytes:

Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)
Chloride	<100	Fluoride	<20	Nitrite	<50	Sulfate	<100
Cyanide	<10	Nitrate	<20	Paraquat (amber only)	<0.4	Sulfide	<30
Diquat (amber only)	<1.0					Sulfite	<1000

Group 2. Glass Sample Containers for use in the analysis of Semivolatiles and Pesticides/PCBs

Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)
Acenaphthene	<5	Acenaphthylene	<5	Anthracene	<5
Benzo(a)anthracene	<5	Benzo(a)pyrene	<5	Benzo(b)fluoranthene	<5
Benzo(k)fluoranthene	<5	Benzo(g,h,i)perylene	<5	Benzoic Acid	<20
Benzyl Alcohol	<5	4-Bromophenyl -phenylether	<5	Butylbenzylphthalate	<5
4-Chloroaniline	<5	4-Chloro -3-methylphenol	<5	bis-(2-Chloroethoxy)methane	<5
bis -(2-Chloroethyl)ether	<5	bis -(2-Chloroisopropyl)ether	<5	2-Chloronaphthalene	<5
2-Chlorophenol	<5	4-Chlorophenyl -phenylether	<5	Chrysene	<5
Di-n-butylphthalate	<5	Di-n-octylphthalate	<5	Dibenzo(a,h)anthracene	<5
Dibenzofuran	<5	1,2-Dichlorobenzene	<5	1,4-Dichlorobenzene	<5
1,3-Dichlorobenzene	<5	3,3'-Dichlorobenzidine	<5	2,4-Dichlorophenol	<5
Diethylphthalate	<5	Dimethylphthalate	<5	2,4-Dinitrotoluene	<5
4,6-Dinitro-2-methylphenol	<20	2,4-Dinitrophenol	<20	Fluoranthene	<5
2,6-Dinitrotoluene	<5	bis -(2-Ethylhexyl)phthalate	<5	Hexachlorobutadiene	<5
Fluorene	<5	Hexachlorobenzene	<5	Indeno(1,2,3 -cd)pyrene	<5
Hexachlorocyclopentadiene	<5	Hexachloroethane	<5	2-Methylphenol	<5
Isophorone	<5	2-Methylnaphthalene	<5	3-Nitroaniline	<20
4-Methylphenol	<5	2-Nitroaniline	<20	N-Nitrosodimethylamine	<5
4-Nitroaniline	<20	N-Nitroso-di-n-propylamine	<5	Nitrobenzene	<5
N-Nitrosodiphenylamine	<5	Naphthlene	<5	Pentachlorophenol	<20
2-Nitrophenol	<5	4-Nitrophenol	<20	Pyrene	<5
Phenanthrene	<5	Phenol	<5	2,4,6 -Trichlorophenol	<5
1,2,4-Trichlorobenzene	<5	2,4,5-Trichlorophenol	<20	Aldrin	<0.01
Azobenzene	<5	Carbazole	<5	Alpha -BHC	<0.01
4,4-DDD	<0.02	Endosulfan II	<0.02	Beta-BHC	<0.01
4,4-DDE	<0.02	Endosulfan Sulfate	<0.02	Delta-BHC	<0.01
4,4-DDT	<0.02	Endrin	<0.02	Gamma-BHC	<0.01
Dieldrin	<0.02	Endrin Aldehyde	<0.02	Heptachlor Epoxide	<0.01
Endosulfan I	<0.01	Heptachlor	<0.01	Alpha-Chlordane	<0.20
Methoxychlor	<0.10	Endrin Ketone	<0.02	Aroclor-1016	<0.20
Gamma-Chlordane	<0.01	Toxaphene	<0.30	Aroclor-1242	<0.20
Aroclor-1221	<0.20	Aroclor-1232	<0.20	Aroclor-1260	<5
Aroclor-1248	<0.20	Aroclor-1254	<0.20	TPH by EPA 1664	<50mg/kg
Aroclor-1262	<0.20	Aroclor-1268	<0.20		
Diesel Range Organics	<0.2mg/L	Oil & Grease	<5.0mg/L		

Group 3. Glass Sample Containers for use in the analysis of Volatiles

Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)	Analyte	Quantitation Limit (µg/L)
Acetone	<5	1,3-Dichloropropane	<1	Benzene	<1
2,2-Dichloropropane	<1	Bromobenzene	<1	1,2-Dichloropropane	<1
Bromodichloromethane	<1	trans -1,3-Dichloropropene	<1	Bromoform	<1
cis-1,3-Dichloropropene	<1	Bromomethane	<1	1,1-Dichloropropene	<1
2-Butanone	<5	Ethylbenzene	<1	tert-Butylbenzene	<1
Hexachlorobutadiene	<1	sec-Butylbenzene	<1	2-Hexanone	<5
n-Butylbenzene	<1	Isopropylbenzene	<1	Carbon Disulfide	<1
p-Isopropyltoluene	<1	Carbon Tetrachloride	<1	4-Methyl -2-pentanone	<5
Chloromethane	<1	Methylene Chloride	<2	Chloroethane	<1
1,1,2,2-Tetrachloroethane	<1	Chloroform	<1	n-Propylbenzene	<1
Dibromochloromethane	<1	Styrene	<1	2 & 4 Chlorotoluene	<1
1,2,3-Trichloropropane	<1	1,2-Dibromo-3-chloropropane	<1	Tetrachloroethene	<1
1,4-Dichlorobenzene	<1	Toluene	<1	1,2-Dibromoethane (EDB)	<1
1,1,1-Trichloroethane	<1	Dibromomethane	<1	1,2,4-Trichlorobenzene	<1
Dichlorodifluoromethane	<1	1,1,2-Trichloroethane	<1	1,3-Dichlorobenzene	<1
1,2,3-Trichloropropane	<1	1,2-Dichlorobenzene	<1	Trichloroethene	<1
trans -1,2-Dichloroethene	<1	Trichlorofluoromethane	<1	1,2-Dichloroethene	<1
Vinyl Acetate	<5	1,1-Dichloroethane	<1	Bromochloromethane	<1
Xylenes (total)	<1	1,3,5-Trimethylbenzene	<1		
Vinyl Chloride	<1	1,1-Dichloroethene	<1		
		1,2,4-Trimethylbenzene	<1		
		cis-1,2-Dichloroethene	<1		

In addition to the above analytes , 40 mL and 60 mL vials are certified for:

Analyte	Quantitation Limit (µg/L)
Total Organic Carbon	<600

If Fisher Scientific can be of any further assistance, please call 800550-4964 and ask for our Technical Service Department.

Approved By:

Robby Ryans
Quality Assurance

Bottle Type: Glass	QA Level: 1	Lot Number: 00085650
Item Number: 02912313	Item Description: 1000mL WM Clear SS Jar	
	Groups 1 and 2 are applicable	