

Certificate Of Analysis

Product Information

Part Number: 181130
Description: AMMONIA RGNT FOR SODIUM ANALYZER, 2 L
Lot Number: YBP1
Expiration Date: No Expiration Date - Suggested use within 2 years from date of receipt by customer/end-user
Method Of Analysis: POTENTIOMETRIC

Actual Analysis

Acceptance Range	Results	Unit Of Measure
28 % TO 30 % AMMONIUM HYDROXIDE	30	% AMMONIUM HYDROXIDE

Traceability

N.I.S.T Chemical	SRM#
NOT APPLICABLE	NA

Certified By: Thermo Fisher Scientific

Water and Lab Products

Quality Assurance

Certificate Date: 02/14/24

244721-001 Rev.K

Water and Lab Products

Australia: (613) 9757-4300 In Australia: (1300) 735-295 China: (86) 21-6865-4588
Germany: (49) 6184-90-6321 India: (91) 22-6716-2261/2247 Japan: (81) 045-453-9175
North America: 1-978-232-6000 Toll Free: 1-800-225-1480 Singapore: (65) 6778-6876

Find out more at thermofisher.com/water