

Certificate of Analysis

Product Information

Part Number : 810001
Description : ROSS ELECTRODE STORAGE SOLUTION, 475ML
Lot Number : SWX3
Expiration Date : 27-MAR-2020
Method of Analysis : POTENTIONMETRIC

Actual Analysis

Acceptance Range	Results	Unit of Measure
3.0 M +/- 0.1M POTASSIUM CHLORIDE	2.97	M POTASSIUM CHLORIDE

Traceability

N.I.S.T. Chemical	SRM#
SODIUM CHLORIDE	2201

Certified By : Thermo Fisher Scientific
Water and Lab products
Quality Assurance
Certificate Date : 27-MAR-2018