

Certificate Of Analysis

Product Information

Part Number:	9100CB
Description:	PH ELECTRODE STORAGE SOLUTION, 19L CUBE
Lot Number:	CR1
Expiration Date:	No Expiration Date - Suggested use within 2 years from date of receipt by customer/end-user
Method Of Analysis:	POTENTIOMETRIC

Actual Analysis

Acceptance Range	Results	Unit Of Measure
3.0 M +/- 0.1 M POTASSIUM CHLORIDE	3.0	M POTASSIUM CHLORIDE
6.75 +/- 0.20 PH UNITS AT 25 C	6.70	PH UNITS AT 25 C

Traceability

N.I.S.T Chemical	SRM#
DISODIUM HYDROGEN PHOSPHATE	186 II g
POTASSIUM DIHYDROGEN PHOSPHATE	186 I g
SODIUM BICARBONATE	191d-I
SODIUM CARBONATE	191d-II

Certified By: Thermo Fisher Scientific

Water and Lab Products

Quality Assurance

Certificate Date: 09/12/24

244721-001 Rev.K

Water and Lab Products

Australia: (613) 9757-4300 In Australia: (1300) 735-295 China: (86) 21-6865-4588
Germany: (49) 6184-90-6321 India: (91) 22-6716-2261/2247 Japan: (81) 045-453-9175
North America: 1-978-232-6000 Toll Free: 1-800-225-1480 Singapore: (65) 6778-6876

Find out more at thermofisher.com/water