

Certificate of Analysis

Product Information

Part Number : 910104
 Description : PH4.01 BUFFER COLOR - PINK, 475ML
 Lot Number : SDZ4
 Expiration Date : 17-Jan-2027
 Method of Analysis : POTENTIONMETRIC

Actual Analysis

Acceptance Range	Results	Unit of Measure
4.01 +/- 0.01 PH UNITS @ 25°C	4.01	PH @ 25°C

Traceability

N.I.S.T. Chemical	SRM#
POTASSIUM DIHYDROGEN PHOSPHATE	186-I-g
DISODIUM HYDROGEN PHOSPHATE	186-II-g
POTASSIUM HYDROGEN PHTHALATE	185I

Certified By : Thermo Fisher Scientific
 Water and Lab Products
 Quality Assurance

Certificate Date : 17-Jan-2025

Find out more at thermofisher.com/water

Contact us today: wlp.techsupport@thermofisher.com

For Research Use Only. Not for use in diagnostic procedures. © 2020 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.