

# Certificate Of Analysis

## Product Information

Part Number: 910107  
Description: PH 7.00 BUFFER, COLOR-YELLOW, 475ML  
Lot Number: CP3  
Expiration Date: 11/30/2026  
Method Of Analysis: POTENTIOMETRIC

## Actual Analysis

Acceptance Range	Results	Unit Of Measure
7.00 +/- 0.01 PH UNITS AT 25 C	7.00	PH UNITS AT 25 C

## Traceability

N.I.S.T Chemical	SRM#
DISODIUM HYDROGEN PHOSPHATE	186 II g
POTASSIUM DIHYDROGEN PHOSPHATE	186 I g
SODIUM BICARBONATE	191d-I
SODIUM CARBONATE	191d-II

Certified By: Thermo Fisher Scientific  
Water and Lab Products  
Quality Assurance

Certificate Date: 11/18/24

244721-001 Rev.K

Water and Lab Products

Australia: (613) 9757-4300 In Australia: (1300) 735-295 China: (86) 21-6865-4588  
Germany: (49) 6184-90-6321 India: (91) 22-6716-2261/2247 Japan: (81) 045-453-9175  
North America: 1-978-232-6000 Toll Free: 1-800-225-1480 Singapore: (65) 6778-6876

Find out more at [thermofisher.com/water](https://thermofisher.com/water)

© 2017 Thermo Fisher Scientific Inc. All rights reserved.  
All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.