

Certificate Of Analysis

Product Information

Part Number: 940907
Description: FLUORIDE 100PPM STANDARD, 475ML
Lot Number: BU1
Expiration Date: 06/30/2025
Method Of Analysis: POTENTIOMETRIC

Actual Analysis

| Acceptance Range | Results | Unit Of Measure |
|-----------------------------------|---------|-----------------|
| 100.0 PPM +/- 0.5 PPM AS FLUORIDE | 100.5 | PPM AS FLUORIDE |

Traceability

| N.I.S.T Chemical | SRM# |
|--------------------|------|
| POTASSIUM FLUORIDE | 2203 |

Certified By: Thermo Fisher Scientific

Water and Lab Products

Quality Assurance

Certificate Date: 06/23/23

244721-001 Rev.K

Water and Lab Products

Australia: (613) 9757-4300 In Australia: (1300) 735-295 China: (86) 21-6865-4588
Germany: (49) 6184-90-6321 India: (91) 22-6716-2261/2247 Japan: (81) 045-453-9175
North America: 1-978-232-6000 Toll Free: 1-800-225-1480 Singapore: (65) 6778-6876

Find out more at thermofisher.com/water

© 2017 Thermo Fisher Scientific Inc. All rights reserved.
All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless
otherwise specified.