

Certificate of Analysis

Product Information:

Product Code	94X377002
Description	Total Chlorine (DPD) Reagent Kit
Source Lot	XB1A1124
Manufacturing Date	30-Nov-2021
Expiration Date	30-Nov-2024

Actual Analysis

Measurement	Acceptance Range	Results
Chlorine determination at 1 mg/l Cl-equivalent (KMnO ₄)	0.200 - 0.245	Pass
pH value	6.15 - 6.5	Pass

Method of Standardization

Chlorine determination by absorbance @ 530nm and determination of pH of reagent in solution. The ambient temperature during testing was 20 °C (+/-2°C).

Approved By

ThermoFisher Scientific
Water and Lab Products
Quality Assurance

This is a computer-generated document. No signature is required

Rev A