

Certificate of Analysis

Product Information:

| | |
|---------------------------|---|
| Product Code | EC4421000BT |
| Description | 1000 PPM @ 25.0 Deg C 442 Conductivity Solution |
| Batch Number | 045/01 |
| Manufacturing Date | 24-Jan-2025 |
| Expiration Date | 24-Jan-2028 |

Actual Analysis

| | | |
|------------------------|-------------------------|----------------|
| Measurement | Acceptance Range | Results |
| Conductivity (@25degC) | 1450 uS/cm +/- 1% | Pass |

Method of Standardization

This solution is standardized using conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurements.

Approved By

ThermoFisher Scientific
Water and Lab Products
Quality Assurance

This is a computer-generated document. No signature is required

Rev A