

Certificate of Analysis

Product Information:

Product Code	EC4421000BT
Description	1000 PPM @ 25.0 Deg C 442 Conductivity Solution
Batch Number	414/01
Manufacturing Date	11-Oct-2024
Expiration Date	11-Oct-2027

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25degC)	1450 uS/cm +/- 1%	Pass

Method of Standardization

This solution is standardized using conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurements.

Approved By

ThermoFisher Scientific
Water and Lab Products
Quality Assurance

This is a computer-generated document. No signature is required

Rev A